LETTERS TO THE EDITOR

Neonatal respiratory distress syndrome

EDITOR,—Professor Southall and colleagues¹ question the comments of the working group on the management of neonatal respiratory distress syndrome regarding parameters of oxygenation.² We write in support of the provisional and practical suggestion of an arterial oxygen saturation (SaO₂) range of 85–93% and suggest that much of the data quoted by Southall *et al* is not applicable to this clinical setting.

Professor Southall and colleagues suggest an upper limit for SaO₂ of 96% and quote four studies in support of this choice. In two of these papers,^{3,4} the majority of the patients were not preterm and were receiving treatment for other conditions, mainly congenital heart disease. These infants constitute a very different population from those suffering from respiratory distress syndrome. Our own study of 477 measurements made in clinical practice using Ohmeda Biox oximeters on a population under 33 weeks' gestation found that if the saturation value was 93% or below the arterial oxygen tension (PaO₂) was never greater than 12 kPa.⁵ To maintain the PaO₂ below 10 kPa, which is the upper limit suggested by the working party, requires a saturation of 92% or less. Certainly, when using an Ohmeda oximeter, an upper limit of 93% seems very reasonable.

When discussing a lower limit for SaO₂, Professor Southall recommends 95%, based on a study of well preterm infants.⁶ These infants bear little relation to the sick, low birthweight infants who, until very recently had experienced a physiological, umbilical venous oxygen tension (PO₂) of approximately 4.5–6.0 kPa.

With regard to the adverse effects of hypoxia, there is no firm evidence which defines the optimal range for any of the indices of oxygenation. There are almost no modern data relating levels of oxygenation to outcome; the papers quoted by Professor Southall regarding hypoxia describe physiological responses to lower levels of oxygenation and not outcome.

Limited by the lack of scientific data, the working party have had to make reasonable and, equally important, practical recommendations. We feel that there is little evidence to support a range of 94 to 96% and in addition would be surprised to find any neonatal unit able to maintain the SaO₂ within such narrow limits. With the marked fluctuations in oxygenation seen in preterm infants receiving intensive care, it would be difficult to make sufficiently frequent adjustments in ventilator settings to comply with these limits. Based on our own data we currently employ the range 86 to 94%, similar to that mentioned in the working party report and look forward to further studies to give us more guidance on this difficult question.

D P COCHRAN
N J SHAW
Regional Neonatal Intensive Care Unit,
Liverpool Maternity Hospital,
Oxford Street,
Liverpool L7 7BN

- 1 Southall DP, Samuels MP, Poets CF. Neonatal respiratory distress syndrome [letter]. *Arch Dis Child* 1993; 69: 330.
- 2 Joint Working Group of the British Association of Perinatal Medicine and the Research Unit of the Royal College of Physicians. Development of audit measures and guidelines for good practice in the management of neonatal respiratory distress syndrome. *Arch Dis Child* 1992; 67: 1221–7.
- 3 Southall DP, Bignall S, Stebbens VA, Alexander JR, Rivers RPA, Lissauer T. Pulse oximeter and transcutaneous arterial oxygen measurements in neonatal and paediatric intensive care. *Arch Dis Child* 1987; 62: 882–8.
- 4 Poets CF, Wilken M, Seidenberg J, Southall DP, von der Hardt H. Reliability of the pulse oximeter in the detection of hyperoxemia. *J Pediatr* 1993; 122: 87–90.
- 5 Cochran DP, Shaw NJ. The use of pulse oximetry in the prevention of hyperoxia in preterm infants. *Pediatric Reviews and Communications* (in press).
- 6 Poets CF, Stebbens VA, Alexander JR, Arrowsmith WA, Salfeld SAW, Southall DP. Arterial oxygen saturation in preterm infants at discharge from the hospital and six weeks later. *J Pediatr* 1992; 120: 447–54.

Professor Southall and colleagues comment:

The upper limit for SaO₂ that we recommend is based on several studies correlating PaO₂ and saturation levels.^{1–3} Although the two papers mentioned by Cochran and Shaw did include older infants,^{1,2} the potential bias would result in an underestimation of the pulse oximeter's sensitivity to hyperoxia – the higher proportion of fetal haemoglobin in preterm infants would result in a higher SaO₂ at any given PaO₂. With regard to avoidance of hyperoxia (PaO₂>13 kPa), the Nellcor N100 has been found to identify hyperoxia with a 95% sensitivity at saturations of 96% and above; the Ohmeda Biox 3700 had a similar sensitivity, but at saturations of 89%.³

The different sensitivity, averaging times, and algorithms of available pulse oximeters must be taken into account when discussing acceptable ranges of SaO₂ in the neonatal population. Until manufacturers standardise their technology, it would be safer to (a) use oximeters that have been validated for use in the neonatal population, with demonstrated sensitivity for the detection of hyperoxia but allowing a range that avoids the risk of hypoxaemia, (b) ensure that when defining acceptable ranges for SaO₂, the range is model specific to avoid inadvertent hypoxia or hyperoxia by using inappropriate limits.

Our recommended lower limit for saturation was originally based on a study of 'well' preterm infants at discharge from hospital. It has now been confirmed in a study of non-distressed preterm neonates investigated during their first week of life.⁴ It does not seem unreasonable to aim to achieve the same range of SaO₂ in 'sick' preterm infants. The argument that the latter group had recently experienced a physiological *intrauterine* PaO₂ of 4.5–6.0 kPa is untenable in the light of the very different stresses of extrauterine life, not least of which are the opening of the pulmonary vascular bed and a significant increase in oxygen consumption.

Finally, larger swings in SaO₂ will be seen in an infant if lower SaO₂ values are maintained. This is because of the shape of the dissociation curve, and the resultant effects of hypoxaemia and airway hypoxia on pulmonary vascular and bronchomotor reactivity time.

- 1 Southall DP, Bignall S, Stebbens VA, Alexander JR, Rivers RPA, Lissauer T. Pulse oximeter and transcutaneous arterial oxygen measurements in neonatal and paediatric intensive care. *Arch Dis Child* 1987; 62: 882–8.
- 2 Poets CF, Wilken M, Seidenberg J, Southall DP, von der Hardt H. Reliability of the pulse

oximeter in the detection of hyperoxemia. *J Pediatr* 1993; 122: 87–90.

- 3 Bucher HU, Fanconi S, Baekert P, Duc G. Hyperoxemia in newborn infants: detection by pulse oximetry. *Pediatrics* 1989; 84: 226–30.
- 4 Richard D, Poets CF, Neale S, Stebbens VA, Alexander JR, Southall DP. Arterial oxygen saturation in preterm neonates without respiratory failure. *J Pediatr* 1993; 123: 963–8.

Blood pressure and blood volume in preterm infants

EDITOR,—We read with interest the article by Bauer *et al* on blood volume and blood pressure in preterm infants.¹ We have looked at a very similar group of infants and assessed the blood volume but with different results.

In 31 infants the mean arterial blood pressure (MABP) was measured by invasive arterial methods. Blood volume was assessed using an indocyanine green dye dilution method described previously in this age group.² The infants had a median gestational age of 26 weeks (range 25–31 weeks) and a median birth weight of 900 g (range 580–1380 g). Readings were all obtained in the first three days of life.

The median blood volume observed in our infants was 85 ml/kg (range 46–131 ml/kg), very similar to both the 83 ml/kg of Bauer *et al* and to previously published data. We found a median MABP of 36 mm Hg (range 18–47). We found no significant relationship, on regression analysis, between observed blood volume and MABP, $p=0.42$ (figure).

Relationship between MABP and blood volume.

Thus, we have not confirmed the third order polynomial relationship described by Bauer *et al*. This may be for a number of reasons, but it may be of note that the majority of the recordings of blood pressure from Bauer's group were by a non-invasive method, that has been previously criticised in the sick very low birthweight infant.³

We would agree with the conclusion of Bauer *et al* that blood pressure assessment alone is a poor measure of absolute hypovolaemia in very low birthweight infants. It is important that more than one cardiovascular parameter is considered in the assessment of the hypotensive infant so that volume expansion and inotropes are used appropriately in this age group.

I M R WRIGHT
Department of Paediatrics and Child Health

S R GOODALL
Department of Biochemistry,
Leeds General Infirmary,
Belmont Grove,
Leeds LS2 9NS

- 1 Bauer K, Linderkamp O, Versmold HT. Systolic blood pressure and blood volume in preterm infants. *Arch Dis Child* 1993; 69: 521–2.
- 2 Anthony MY, Goodall SR, Papouli M, Levene MI. Measurement of plasma volume in neonates. *Arch Dis Child* 1992; 67: 36–40.
- 3 Weindling AM. Blood pressure monitoring in the newborn. *Arch Dis Child* 1989; 64: 444–7.