ORIGINAL ARTICLE

Monitoring of end tidal carbon dioxide and transcutaneous carbon dioxide during neonatal transport

D G Tingay, M J Stewart, C J Morley

See end of article for authors' affiliations

Arch Dis Child Fetal Neonatal Ed 2005;90:F523-F526. doi: 10.1136/adc.2004.064717

Correspondence to:
Dr Tingay, Department of
Neonatology, Royal
Children's Hospital,
Flemington Rd, Parkville,
Victoria 3052, Australia;
david.tingay@rch.org.au

Accepted 22 April 2005
Published Online First
29 April 2005

Objective: To assess the accuracy of measurements of end tidal carbon dioxide (CO_2) during neonatal transport compared with arterial and transcutaneous measurements.

Design: Paired end tidal and transcutaneous CO_2 recordings were taken frequently during road transport of 21 ventilated neonates. The first paired CO_2 values were compared with an arterial blood gas. The differences between arterial CO_2 (PaCO_2), transcutaneous CO_2 (TcPCO_2), and end tidal CO_2 (PetCO_2) were analysed. The Bland-Altman method was used to assess bias and repeatability.

Results: PetCO_2 correlated strongly with PaCO_2 and TcPCO_2 . However, PetCO_2 underestimated PaCO_2 at a clinically unacceptable level (mean (SD) 1.1 (0.70) kPa) and did not trend reliably over time within individual subjects. The PetCO_2 bias was independent of PaCO_2 and severity of lung disease.

Conclusions: PetCO_2 had an unacceptable under-recording bias. TcPCO_2 should currently be considered the preferred method of non-invasive CO_2 monitoring for neonatal transport.

Continuous non-invasive carbon dioxide (CO_2) monitoring has become an important bedside tool in neonatal intensive care. Transported sick neonates should receive full intensive care, but arterial blood gas monitoring is not possible. Assessing the efficacy of ventilation during neonatal transport is challenging. Continuous non-invasive CO_2 monitoring has been shown to increase the likelihood of the patient arriving at the receiving hospital with a normal pH and partial pressure of CO_2 (PaCO_2).¹

Transcutaneous CO_2 monitoring is the most commonly used non-invasive CO_2 monitoring system in neonatal intensive care and has been shown to accurately predict PaCO_2 and monitor CO_2 trends.^{1,2} Calibrated transcutaneous partial pressure of carbon dioxide (TcPCO_2) has been shown to reliably approximate PaCO_2 during neonatal transport and has been recommended as an alternative to frequent PaCO_2 measurements.¹ However, TcPCO_2 devices are difficult to use,^{3,4} bulky, and weigh between 2 and 6 kg, thus limiting their use during neonatal transport.

End tidal CO_2 (PetCO_2) monitors are lightweight and may indirectly monitor PaCO_2 .⁵⁻⁸ Hence, PetCO_2 may be more useful during transportation than TcPCO_2 monitoring. Studies of PetCO_2 monitoring in newborn infants have had mixed results, primarily because of the effects of ventilation perfusion mismatching on PetCO_2 , failure to reach an expiratory plateau during rapid respiratory rates, and the technical limitations of PetCO_2 devices to interpret CO_2 in small tidal volume states.^{2,5,9-12} Recent technological advances in PetCO_2 monitoring, such as smaller sample volumes and sample cells calibrated to neonatal tidal volumes, have attempted to overcome the limitations.¹³ Some authors advocate PetCO_2 as an acceptable method of approximation of PaCO_2 trends in newborn infants.^{10,14-16}

The Newborn Emergency Transport Service of Victoria (NETS) is the largest neonatal transport service in Australasia. More than 900 infants a year are transported, with approximately one third ventilated. Monitoring of TcPCO_2 and oxygen saturation have been standard practice for five years to indicate ventilation adequacy during transport, and previous unpublished data have shown a close correlation between TcPCO_2 and PaCO_2 .

Arterial blood gases and TcPCO_2 are commonly used to monitor ventilation. The aim of this study was to assess the accuracy and reliability of PetCO_2 monitoring during neonatal transport.

METHODS

Ventilated infants requiring road transport to a level 3 neonatal intensive care unit during March to August 2002 were recruited if the paediatrician involved in the transport was specifically trained to use both PetCO_2 and TcPCO_2 monitors, an arterial catheter was being used, endotracheal tube position could be confirmed by chest radiograph before transport, and both TcPCO_2 and PetCO_2 monitoring could be started before the first arterial blood gas was measured by the NETS team. Because of the effects of barometric pressure on PetCO_2 , infants transported by air were not studied.⁵ Informed parental consent was obtained for each infant before transport.

Infants were not studied if they were older than 28 days, had a capillary refill time of greater than two seconds, or TcPCO_2 or PetCO_2 readings could not be made or were lost during transport.

TcPCO_2 was measured using the Microgas 7650 system (weight 5.6 kg) with Combi.M sensor 82 (Linde, Basel, Switzerland) applied to the skin of the anterior chest or abdomen. The manufacturers report that the Combi.M sensor 82, once calibrated, will remain accurate for up to four hours at one site. PetCO_2 was measured using a side stream end tidal analyser specifically designed for neonatal use (the Agilent Microstream system; Agilent Technologies, Andover, Massachusetts, USA); a result was the highest of five consecutive measurements.¹³ Arterial blood gases were analysed with the i-STAT portable clinical analyser (i-STAT Corporation, East Windsor, New Jersey, USA). Infants were ventilated using the Hoekloos Infant ventilator Mark 3

Abbreviations: PaCO_2 , arterial partial pressure of carbon dioxide; TcPCO_2 , transcutaneous partial pressure of carbon dioxide; PetCO_2 , end tidal partial pressure of carbon dioxide; NETS, Newborn Emergency Transport Service (Victoria); $\text{PAO}_2/\text{PaO}_2$ ratio, alveolar-arterial oxygen tension ratio

Table 1 Characteristics of the 21 subjects enrolled in study

	Median	Range
Gestational age (weeks)	35	26–40
Birth weight (g)	2260	930–4600
Age at enrolment (hours)	4.8	1.8–61.2
Transportation time (minutes)	65	20–180
	Mean (SD)	Range
pH	7.32 (0.12)	7.1–7.55
FiO ₂	0.52 (0.24)	0.21–1.0
PAO ₂ /PaO ₂ ratio	0.85 (1.3)	0.03–5.9
Primary diagnosis	Number	
Respiratory failure	15	
Cyanotic heart disease	2	
Persistent pulmonary hypertension of the newborn	1	
Severe anaemia	1	
Birth asphyxia	1	
Multiple congenital abnormalities	1	

FiO₂, Inspired oxygen fraction; PAO₂/PaO₂ ratio, alveolar-arterial oxygen tension ratio.

(Hoekloos, Amsterdam, Netherlands). The Australian Therapeutics Goods Administration has approved both devices for use in newborn infants. A specialist neonatal transport nurse and neonatal paediatrician escorted all infants.

After calibration of the TcPco₂ and Petco₂ monitors, paired CO₂ measurements were recorded every 20 minutes, starting at stabilisation and continuing throughout the transport. The initial recordings were calibrated with a simultaneous Paco₂. The NETS team was not blinded to the TcPco₂ or Petco₂ values; any ventilator changes were based on the TcPco₂ or Paco₂ values.

The severity of each baby's lung disease was determined by calculating the alveolar to arterial oxygen tension ratio (PAO₂/Pao₂ ratio) where PAO₂ = (Barometric pressure – 47) × (FiO₂ – Pao₂). Severe lung disease was defined as a PAO₂/Pao₂ ratio <0.3. A PAO₂/Pao₂ ratio of <0.3 has been associated with less precision of Petco₂ measurements to estimate Paco₂.¹⁵

The parents of all infants enrolled in the study provided written and signed informed consent for their infants to be transported by NETS and this involved specific consent to the use of all devices used in the study. This study was discussed with the Royal Women's Hospital Ethics in Human Research Committee. It was decided that formal ethics approval was not required as the above written informed consent adequately informed the parents and addressed the ethical issues of the study.

Statistical analysis

The differences between Paco₂, TcPco₂, and Petco₂ (expressed as P_(a-Tc)CO₂, P_(a-Et)CO₂, and P_(Tc-Et)CO₂ respectively) were analysed using a Student's paired *t* test, and their correlations were calculated. The Bland-Altman technique

Figure 1 Bland-Altman plot of the difference between Paco₂ and Petco₂ (P_(a-Et)CO₂) against average CO₂. **Figure 2** Bland-Altman plot of the difference between Paco₂ and TcPco₂ (P_(a-Tc)CO₂) against average CO₂.

was used to assess agreement and repeatability.¹⁷ A bias of less than ± 0.7 kPa was considered clinically acceptable. Intrasubject P_(Tc-Et)CO₂ variability over time was calculated.

RESULTS

Twenty six infants were enrolled, but five were excluded because the Petco₂ could not be continuously measured in three, both TcPco₂ and Petco₂ could not be measured in another, and in the fifth infant the initial blood gas was venous. Table 1 summarises the characteristics of the 21 infants. A total of 21 P_(a-Tc)CO₂ and P_(a-Et)CO₂ differences and 82 P_(Tc-Et)CO₂ differences (median recordings per subject 4.0 (range 2–10)) were calculated.

There was a linear relation between Petco₂, Paco₂, and TcPco₂. However, Petco₂ underestimated Paco₂ by an average of 1.04 kPa (table 2, fig 1). Only 48% of Petco₂ recordings were within 1.0 kPa of the paired Paco₂. The bias of the Petco₂ values was independent of the Paco₂.

TcPco₂ was closely related to Paco₂, with no significant difference between the two measurements (table 2). Two thirds of TcPco₂ readings were within 0.7 kPa of the Paco₂,

Table 2 A comparison of CO₂ (kPa) measured in three different ways

	n	Mean (SD)	95% CI	p Value
P _(a-Tc) CO ₂	21	–0.13 (0.71)	–0.46 to 0.19	0.4
P _(a-Et) CO ₂	21	1.04 (0.98)	0.59 to 1.49	<0.001
P _(Tc-Et) CO ₂	82	–0.07 (0.84)	–0.26 to 0.11	0.43

and 81% of TcP_{CO_2} readings were within 1 kPa of the paired P_{aCO_2} . There was no significant change in the difference between TcP_{CO_2} and P_{aCO_2} as the CO_2 level changed (fig 2).

When the initial TcP_{CO_2} and P_{etCO_2} values for each subject were calibrated to the original P_{aCO_2} , there was a closer relation between P_{etCO_2} and TcP_{CO_2} : 64% of P_{etCO_2} values were within 0.7 kPa of the paired TcP_{CO_2} value (fig 3). Although the $P_{(Tc-Et)CO_2}$ difference was not significant, the variability, as demonstrated by the Bland-Altman plot, was large (table 2, fig 3).

There was no significant relation between P_{etCO_2} accuracy and severity of lung disease (table 3), although there was a non-significant trend towards P_{etCO_2} values being more likely to reflect either P_{aCO_2} or TcP_{CO_2} in infants with a PAO_2/P_{aO_2} ratio >0.3 . Muscle relaxation did not alter the reliability of P_{etCO_2} to trend with TcP_{CO_2} .

DISCUSSION

This study shows that, in neonates requiring ventilation during transport, TcP_{CO_2} monitoring more accurately reflected P_{aCO_2} than P_{etCO_2} monitoring. Furthermore, P_{etCO_2} monitoring should be used with caution. Both P_{etCO_2} and TcP_{CO_2} were linearly related to P_{aCO_2} and each other. However, a linear relation alone (or correlation coefficients—the method used in many of the previous reports) does not adequately describe the agreement between two clinical measurement techniques.^{2 10 18} Assessing agreement between two methods of clinical measurement is complex. The method described by Bland and Altman is a more informative technique for assessing agreement, reliability, and repeatability, and allows interpretation within a clinical context.¹⁷ With the use of this technique, P_{etCO_2} was neither as precise nor reliable a method of assessing P_{aCO_2} during the transport of ventilated neonates, whereas TcP_{CO_2} provided a more reliable method. The degree of bias demonstrated between P_{etCO_2} and P_{aCO_2} (1.04 kPa) is clinically unacceptable.

Most of the infants in this study had mechanical ventilation instigated by the transport team; knowledge of any changes in the CO_2 is essential for safe delivery of ventilation. Frequent P_{aCO_2} measurements are not practical during neonatal transport; a reliable non-invasive indicator of P_{aCO_2} is essential. Calibrated TcP_{CO_2} is an acceptable surrogate for P_{aCO_2} trends over time. Transcutaneous gas monitoring is an established and validated practice in neonatology.³ Newborn infants are particularly suited to transcutaneous monitoring because of their thin skin. Although proper use is dependent on appropriate training and placement, the only practical limitations are skin perfusion (which may be altered by vasoconstrictive agents,

Figure 3 Bland-Altman plot of the difference between TcP_{CO_2} and P_{etCO_2} ($P_{(Tc-Et)CO_2}$) against average CO_2 .

What is already known on this topic

- TcP_{CO_2} has been shown to be an accurate and reliable method of indicating P_{aCO_2} in neonates receiving intensive care
- Although measurement of P_{etCO_2} can also indicate endotracheal tube position, in previous studies the ability to accurately reflect P_{aCO_2} has been variable

What this study adds

- This study shows that TcP_{CO_2} accurately reflects P_{aCO_2} during neonatal transport, whereas P_{etCO_2} underestimates P_{aCO_2} by about 1.0 kPa, a clinically unacceptable difference
- P_{etCO_2} was also unable to reliably reflect TcP_{CO_2} over time, therefore this study supports the use of TcP_{CO_2} as the preferred method of non-invasive CO_2 monitoring during neonatal transport

hypovolaemia, and oedema) and the temperature produced by the device. The response time of TcP_{CO_2} is too slow (30–50 seconds) to allow monitoring of the respiratory pattern.¹⁹ TcP_{CO_2} monitoring in neonatal transport has previously been evaluated and shown to result in improved ventilation on arrival at the receiving institution.^{1 20}

Many authors have reported a good correlation between P_{etCO_2} , TcP_{CO_2} , and P_{aCO_2} in newborn infants, but in only three studies that evaluated P_{etCO_2} was the relation assessed using the Bland-Altman technique.^{14 15 21} Rozycki *et al*¹⁴ described a mean (SD) $P_{(a-Et)CO_2}$ bias of 0.92 (0.92) kPa in 45 newborn infants receiving mechanical ventilation, with only 36.9% of P_{etCO_2} values falling within 0.67 kPa of the P_{aCO_2} . The authors concluded that despite the significant bias, P_{etCO_2} provided a reliable estimate of P_{aCO_2} trends. A similar mean $P_{(a-Et)CO_2}$ difference of 0.91 (0.68) kPa was reported by Tobias and Meyer²¹ in 25 infants and toddlers (up to 48 months of age) receiving mechanical ventilation for respiratory failure; the $P_{(a-Tc)CO_2}$ difference in this study was 0.31 (0.18) kPa. Sivan *et al*¹⁵ obtained a clinically acceptable $P_{(a-Et)CO_2}$ result, with a mean difference of 0.45 (0.88) kPa in a study involving 134 children (aged 2 days to 16 years) receiving mechanical ventilation. The mean $P_{(a-Tc)CO_2}$ in this group was -0.17 (0.96) kPa. The $P_{(a-Tc)CO_2}$ bias was related to skin perfusion but remained clinically acceptable. Primary diagnosis was not described in this study, nor was the proportion of the population who were newborn infants, making inference to the neonatal population difficult. Sivan and colleagues concluded that the degree of the $P_{(a-Et)CO_2}$ bias was reduced in children with mild lung disease, as defined by a PAO_2/P_{aO_2} ratio of >0.3 . In the cohort with severe lung disease, the mean $P_{(a-Et)CO_2}$ 1.04 (0.97) kPa was similar to our study.

Parenchymal lung disease with ventilation perfusion (V/Q) mismatching and a $PAO_2/P_{aO_2} <0.3$ is a feature of most causes of neonatal respiratory failure. During our study, only two infants did not require oxygen, and nearly all had parenchymal lung disease. Our study was not designed to assess the relation between degree of lung disease and P_{etCO_2} accuracy.

P_{etCO_2} monitoring has been validated in adult ventilated patients and healthy anaesthetised infants, but the infants in

Table 3 Relation between PetCO₂ values and severity of lung disease

	Severe (n = 12)			Mild-moderate (n = 8)		
	Mean (SD)	95% CI	p Value	Mean (SD)	95% CI	p Value
P _(a-Et) CO ₂	1.21 (0.76)	0.87 to 1.88	<0.001	0.99 (1.16)	-0.61 to 1.37	0.013

All CO₂ values in kPa. Severe lung disease, PAO₂/PaO₂ ratio <0.3; mild-moderate lung disease, PAO₂/PaO₂ ratio ≥0.3.

our study had respiratory failure.^{10–18} PetCO₂ is dependent on alveolar CO₂ (PACO₂) and the site of sampling. Non-uniform alveoli CO₂ emptying patterns in patients with large ventilation-perfusion mismatching result in PACO₂ underestimating PaCO₂.^{5–22}

Technical limitations of end tidal analysis in patients with high rate, low tidal volume breathing would have contributed to the difference between PetCO₂ and PACO₂. To account for the fresh inhaled gas admixture during proximal PetCO₂ sampling, a minimum sampling flow rate of 150 ml/min is required.⁵ The end tidal analyser used in our study sampled at 50 ml/min. Despite manufacturer assurances, this may have had an impact on our results. The response time of end tidal analysers must be less than the respiratory cycle. The response time of the end tidal analyser used was 190 milliseconds, which is adequate for the ventilation rates used during the study, although at high respiratory rates with a short expiratory time, all exhaled alveolar gas would not have migrated to a proximal end tidal sampling site on completion of each respiratory cycle.⁵

The relation between TcPCO₂ and PetCO₂ was not constant over time within individuals, even when both values were adjusted to PACO₂. In our opinion PetCO₂ monitoring cannot be used to reliably monitor trends in PACO₂ over time in newborn infants with lung disease.

Despite our findings, PetCO₂ monitoring may offer some benefits over TcPCO₂ monitoring. Primarily the ability to rapidly and reliably confirm endotracheal tube position within the trachea, with either a capnograph or colorimetric end tidal CO₂ indicator, is of great benefit within the noisy environment of neonatal transport.⁷ This study did not aim to assess the ability of PetCO₂ or TcPCO₂ to indicate endotracheal tube position. Inadvertent extubation is not a common occurrence in our transport population and did not occur in any of the neonates involved in this study. Further study is required to determine the role of PetCO₂ in ensuring the endotracheal tube position during transport.

CONCLUSIONS

Owing to the bias of about -1 kPa and lack of consistency in measuring PACO₂ over time, PetCO₂ cannot be recommended during neonatal transport to monitor ventilation. TcPCO₂ monitoring was generally more precise, reliable, and agreed with PACO₂. TcPCO₂ monitoring is the preferred method of non-invasive CO₂ monitoring during neonatal transport.

Authors' affiliations

D G Tingay, M J Stewart, C J Morley, Neonatal Emergency Transport Service (Victoria), Royal Women's Hospital, Carlton, Victoria 3053, Australia

D G Tingay, M J Stewart, Department of Neonatology, Royal Children's Hospital, Parkville, Victoria 3052, Australia

M J Stewart, C J Morley, Department of Neonatology, Royal Women's Hospital

D G Tingay, C J Morley, Murdoch Childrens Research Institute, Parkville, Victoria 3052, Australia

Competing interests: none declared

REFERENCES

- O'Connor TA, Grueber R. Transcutaneous measurement of carbon dioxide tension during long-distance transport of neonates receiving mechanical ventilation. *J Perinatol* 1998;**18**:189–92.
- McEvedy BA, McLeod ME, Mulera M, et al. End-tidal, transcutaneous, and arterial pCO₂ measurements in critically ill neonates: a comparative study. *Anesthesiology* 1988;**69**:112–16.
- Huch A. Transcutaneous blood gas monitoring. *Acta Anaesthesiol Scand Suppl* 1995;**107**:87–90.
- Poets CF, Martin RJ. Noninvasive determinants of blood gases. In: Stocks J, Sly P, Tepper RS, et al, eds. *Infant respiratory function testing*. New York: Wiley-Liss, 1996:411–44.
- Bhavani-Shankar K, Moseley H, Kumar AY, et al. Capnometry and anaesthesia. *Can J Anaesth* 1992;**39**:617–32.
- Shankar KB, Moseley H, Kumar Y. Relationship between PaCO₂-PetCO₂ gradient and physiological dead space. *Can J Anaesth* 1991;**38**:1072–4.
- Bhende MS, Karr VA, Wiltse DC, et al. Evaluation of a portable infrared end-tidal carbon dioxide monitor during pediatric interhospital transport. *Pediatrics* 1995;**95**:875–8.
- Bhende MS, Thompson AE, Orr RA. Utility of an end-tidal carbon dioxide detector during stabilization and transport of critically ill children. *Pediatrics* 1992;**89**:1042–4.
- Watkins AM, Weindling AM. Monitoring of end tidal CO₂ in neonatal intensive care. *Arch Dis Child* 1987;**62**:837–9.
- Badgwell JM, Heavner JE. End-tidal carbon dioxide pressure in neonates and infants measured by aspiration and flow-through capnography. *J Clin Monit* 1991;**7**:285–8.
- Hillier SC, Badgwell JM, McLeod ME, et al. Accuracy of end-tidal PCO₂ measurements using a sidestream capnometer in infants and children ventilated with the Sechrist infant ventilator. *Can J Anaesth* 1990;**37**:318–21.
- Rich GF, Sconzo JM. Continuous end-tidal CO₂ sampling within the proximal endotracheal tube estimates arterial CO₂ tension in infants. *Can J Anaesth* 1991;**3**:201–3.
- Agilent M3015A Microstream CO₂ measurement server extension data sheet. Andover, MA: Agilent Technologies Inc, 2000.
- Rozycki HJ, Sysyn GD, Marshall MK, et al. Mainstream end-tidal carbon dioxide monitoring in the neonatal intensive care unit. *Pediatrics* 1998;**101**:648–53.
- Sivan Y, Eldadah MK, Cheah TE, et al. Estimation of arterial carbon dioxide by end-tidal and transcutaneous PCO₂ measurements in ventilated children. *Pediatr Pulmonol* 1992;**12**:153–7.
- Rich GF, Sullivan MP, Adams JM. Is distal sampling of end-tidal CO₂ necessary in small subjects? *Anesthesiology* 1990;**73**:265–8.
- Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 1986;**1**:307–10.
- McDonald MJ, Montgomery VL, Cerrito PB, et al. Comparison of end-tidal CO₂ and PaCO₂ in children receiving mechanical ventilation. *Pediatr Crit Care Med* 2002;**3**:244–9.
- Schibler A, Frey U. Role of lung function testing in the management of mechanically ventilated infants. *Arch Dis Child Fetal Neonatal Ed* 2002;**87**:F7–10.
- Jacob J, Rose D, Stilson M, et al. Transcutaneous carbon dioxide monitoring during neonatal transport. *Crit Care Med* 1986;**14**:1050–2.
- Tobias JD, Meyer DJ. Noninvasive monitoring of carbon dioxide during respiratory failure in toddlers and infants: end-tidal versus transcutaneous carbon dioxide. *Anesth Analg* 1997;**85**:55–8.
- West J. *Pulmonary pathophysiology: the essentials*, 6th ed. Philadelphia: Lippincott Williams and Wilkins, 2003.